



Ref. Certif. No.

**US-45481-M1-UL**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product	Progressa Hospital Bed
Name and address of the applicant	Hill-Rom Inc 1069 State Rte 46 E Batesville, IN 47006 United States
Name and address of the manufacturer	Hill-Rom Inc 1069 State Rte 46 E Batesville, IN 47006 United States
Name and address of the factory	Hill-Rom Inc 1069 State Rte 46 E Batesville, IN 47006 United States
Note: When more than one factory, please report on page 2	<input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Ratings and principal characteristics	100 VAC, 50/60 Hz, 10 A <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	P750X, Where "X" can be "0" or "1" <input type="checkbox"/> <a href="#">Additional Information on page 2</a>
Additional information (if necessary may also be reported on page 2)	Based on a demonstrated RM file evidence usage of IEC 60601-1:2005+AMD1:2012+AMD2:2020 standard with IEC 60601-2-52:2009+AMD1:2015 was approved, and consequentially TRF IEC60601_1U was used for completion of CBTR. The report was revised to include technical modifications. <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-6:2010, IEC 60601-1-6:2010/AMD1:2013, IEC 60601-1-6:2010/AMD2:2020, IEC 60601-2-52:2009, IEC 60601-2-52:2009/AMD1:2015
As shown in the Test Report Ref. No. which forms part of this Certificate	E351608-D1016-2/A1/C1-CB issued on 2026-01-02

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2026-01-07  
Original Issue Date: 2025-08-20

Signature: Mauricio Avila



Ref. Certif. No.

**US-45481-M1-UL**

**Factory(ies):**

HILL-ROM S A S  
Z I DU TALHOUE, BOITE POSTALE 14 PLUVIGNER 56330  
France

HILL-ROM DE MEXICO S DE RL DE CV  
AV DEL TELEFONO 200, HUINALA APODACA, Nuevo Leon 66640  
Mexico

**Additional Ratings:**

- 110 VAC, 50/60 Hz, 10 A
- 115 VAC, 50/60 Hz, 10 A
- 120 VAC, 50/60 Hz, 10 A
- 120 VAC, 60 Hz, 10 A (US)
- 127 VAC, 50/60 Hz, 10 A
- 220 VAC, 50/60 Hz, 6 A
- 230 VAC, 50/60 Hz, 6 A
- 240 VAC, 50/60 Hz, 6 A

**Summary of Modifications:**

- Update the List of Critical Components table to remove Power Cords for Continental Europe, UK/Ireland, Swiss, Italy/Chile, Danish, Australian/New Zealand, China, Japan, Israel, Argentina, and Brazil.
- Updated International Power Cord Technical Data to cover the removed individual country power cords. And added Letter of Assurance for International Power Cord

**Additional information (if necessary)**



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2026-01-07  
Original Issue Date: 2025-08-20

Signature:

Mauricio Avila